



User's Manual



CYFRA 21-1 CLIA

IVD

REF

CLA-473 I



96 Wells



Legal Manufacturer:

DRG

DRG Instruments GmbH, Germany

Division of DRG International, Inc

Frauenbergstr. 18, D-35039 Marburg

Telefon: +49 (0)6421-17000 Fax: +49-(0)6421-1700 50

Internet: www.drg-diagnostics.de

E-mail: drg@drg-diagnostics.de

Distributed by:



Contents / Inhaltsverzeichnis

| | | |
|----|--|----|
| 1 | INTRODUCTION | 2 |
| 2 | PRINCIPLE OF THE TEST | 2 |
| 3 | WARNINGS AND PRECAUTIONS | 2 |
| 4 | REAGENTS | 4 |
| 5 | SPECIMEN COLLECTION AND PREPARATION | 6 |
| 6 | ASSAY PROCEDURE | 6 |
| 7 | EXPECTED NORMAL VALUES | 8 |
| 8 | QUALITY CONTROL | 8 |
| 9 | PERFORMANCE CHARACTERISTICS | 8 |
| 10 | LIMITATIONS OF USE | 9 |
| 11 | LEGAL ASPECTS | 10 |
| 12 | REFERENCES | 10 |

| | | |
|----|-----------------------------|----|
| 1 | EINLEITUNG | 11 |
| 2 | TESTPRINZIP | 11 |
| 3 | VORSICHTSMAßNAHMEN | 11 |
| 4 | BESTANDTEILE DES KITS | 12 |
| 5 | PROBENVORBEREITUNG | 14 |
| 6 | TESTDURCHFÜHRUNG | 14 |
| 7 | ERWARTETE WERTE | 16 |
| 8 | QUALITÄTS-KONTROLLE | 16 |
| 9 | ASSAY CHARACTERISTIKA | 16 |
| 10 | GRENZEN DES TESTS | 17 |
| 11 | RECHTLICHE GRUNDLAGEN | 17 |
| 12 | REFERENZEN | 17 |

| | | |
|--|-----------------------------------|----|
| | SYMBOLS USED WITH DRG CLIAS | 18 |
|--|-----------------------------------|----|

1 INTRODUCTION

1.1 Intended Use

The **DRG CYFRA 21-1 CLIA** is a chemiluminescence immunoassay for the quantitative *in vitro diagnostic* measurement of CYFRA 21-1 in serum and heparin plasma

1.2 Summary and Explanation

CYFRA 21-1 is a fragment of cytokeratin 19. Although expressed in all body tissues its major occurrence is in the lung, particularly in lung cancer tissues. The major diagnostic importance of CYFRA 21-1 as a tumor marker is in differential diagnosis, prognosis, and aftercare of non-small-cell lung cancer (NSCLC) patients. Additionally, CYFRA 21-1 has been described as a tumor marker for the monitoring of bladder cancer.

The DRG CYFRA 21-1 CLIA uses the two mouse monoclonal antibodies KS19.1 and BM19.21 to determine cytokeratin 19 fragments.

2 PRINCIPLE OF THE TEST

The DRG CYFRA 21-1 CLIA is a chemiluminescence immunoassay (CLIA), based on the sandwich principle. The microtiter wells are coated with a monoclonal [mouse] antibody directed towards a unique antigenic site on the CYFRA 21-1 molecule.

An aliquot of patient sample containing endogenous CYFRA 21-1 is incubated in the coated well with enzyme conjugate, which is an anti CYFRA 21-1 antibody conjugated with horseradish peroxidase. After incubation the unbound conjugate is washed off.

The amount of bound peroxidase is proportional to the concentration of CYFRA 21-1 in the sample.

After addition of the substrate solution, the intensity of emitted light is proportional to the concentration of CYFRA 21-1 in the patient sample.

3 WARNINGS AND PRECAUTIONS

1. This kit is for *in vitro* diagnostic use only. For professional use only.
2. All reagents of this test kit which contain human serum or plasma have been tested and confirmed negative for HIV I/II, HBsAg and HCV by FDA approved procedures. All reagents, however, should be treated as potential biohazards in use and for disposal.
3. Before starting the assay, read the instructions completely and carefully. Use the valid version of the package insert provided with the kit. Be sure that everything is understood.
4. The microplate contains snap-off strips. Unused wells must be stored at 2 °C to 8 °C in the sealed foil pouch and used in the frame provided.
5. Pipetting of samples and reagents must be done as quickly as possible and in the same sequence for each step.
6. Use reservoirs only for single reagents. This especially applies to the substrate reservoirs. Using a reservoir for dispensing a substrate solution that had previously been used for the conjugate solution may turn solution colored. Do not pour reagents back into vials as reagent contamination may occur.
7. Mix the contents of the microplate wells thoroughly to ensure good test results. Do not reuse microwells.
8. Do not let wells dry during assay; add reagents immediately after completing the rinsing steps.
9. Allow the reagents to reach room temperature (21-26°C) before starting the test. Temperature will affect the absorbance readings of the assay. However, values for the patient samples will not be affected.
10. Never pipet by mouth and avoid contact of reagents and specimens with skin and mucous membranes.
11. Do not smoke, eat, drink or apply cosmetics in areas where specimens or kit reagents are handled.
12. Wear disposable latex gloves when handling specimens and reagents. Microbial contamination of reagents or specimens may give false results.
13. Handling should be done in accordance with the procedures defined by an appropriate national biohazard safety guideline or regulation.
14. Do not use reagents beyond expiry date as shown on the kit labels.
15. All indicated volumes have to be performed according to the protocol. Optimal test results are only obtained when using calibrated pipettes and microtiter plate luminometer.
16. The luminescence substrate reagents (*Reagent A* and *Reagent B*) are sensitive to light and should be stored in the original dark bottle away from direct sunlight.

17. Do not mix or use components from kits with different lot numbers. It is advised not to exchange wells of different plates even of the same lot. The kits may have been shipped or stored under different conditions and the binding characteristics of the plates may result slightly different.
18. Some reagents contain Proclin 300, BND and/or MIT as preservatives. In case of contact with eyes or skin, flush immediately with water.
19. Chemicals and prepared or used reagents have to be treated as hazardous waste according to the national biohazard safety guideline or regulation.
20. For information on hazardous substances included in the kit please refer to Material Safety Data Sheets. Material Safety Data Sheets for this product are available upon request directly from DRG.

4 REAGENTS

4.1 Reagents provided

1. **Microtiterwells**, 12x8 strips, 96 wells (break apart);
Wells coated with a anti-CYFRA 21-1 antibody (monoclonal).
2. **Standard (Standard 0-4)**, 5 vials (lyophilized), 1.0 mL;
Concentrations: 0 – 3 – 10 – 25 – 50 ng/mL
see "Reagent Preparation";
* contain 0.03% Proclin 300, 0.015% BND and 0.010% MIT as preservatives.
3. **Control**, 1 vial (lyophilized), 1.0 mL;
see "Reagent Preparation".
For control values and ranges please refer to vial label or QC-Datasheet.
* contain 0.03% Proclin 300, 0.015% BND and 0.010% MIT as preservatives.
4. **Sample Diluent**, 1 vial, 3 mL, ready to use;
* contain 0.03% Proclin 300, 0.015% BND and 0.010% MIT as preservatives.
5. **Enzyme Conjugate 21X concentrate**, 1 vial, 0.5 mL,
anti-CYFRA 21-1 antibody conjugated to horseradish peroxidase
see "Reagent Preparation"
* contains 0.03% Proclin 300, 0.015% BND and 0.010% MIT as preservatives.
6. **Conjugate Diluent**, 1 vial, 7 mL, ready to use.
* contain 0.03% Proclin 300, 0.015% BND and 0.010% MIT as preservative.
7. **Substrate Solution**,
Reagent A, 1 vial, 2.0 mL, *Note: light sensitive!*
Reagent B, 1 vial, 2.0 mL, *Note: light sensitive!*
Reagent C, 1 vial, 5.0 mL
see "Reagent Preparation".
8. **Wash Solution**, 1 vial, 30 mL (40X concentrated);
see „ Reagent Preparation “.

- * BND = 5-bromo-5-nitro-1,3-dioxane
MIT = 2-methyl-2H-isothiazol-3-one

Note: Additional *Sample Diluent* for sample dilution is available upon request.

4.2 Materials required but not provided

- A microtiter plate luminometer.
- Calibrated variable precision micropipettes.
- Absorbent paper.
- Distilled or deionized water
- Timer
- Semi logarithmic graph paper or software for data reduction

4.3 Storage Conditions

When stored at 2°C to 8°C unopened reagents will retain reactivity until expiration date. Do not use reagents beyond this date.

Opened reagents must be stored at 2°C to 8°C. Microtiter wells must be stored at 2°C to 8°C. Once the foil bag has been opened, care should be taken to close it tightly again.

Note: The reconstituted standards and controls are stable for at least 4 weeks at 2°C to 8°C. For longer storage freeze at -20°C.

4.4 Reagent Preparation

Allow all reagents and required number of strips to reach room temperature prior to use.

Standards

Reconstitute the lyophilized contents of the standard vial with 1 mL Aqua dest.

Note: The reconstituted standards are stable for at least 4 weeks at 2°C to 8°C.

For longer storage freeze at -20°C.

Control

Reconstitute the lyophilized content with 1 mL Aqua dest. and let stand for 10 minutes in minimum. Mix the control several times before use.

Note: The reconstituted control is stable for at least 4 weeks at 2°C to 8°C. For longer storage freeze at -20°C.

Wash Solution

Add deionized water to the 40X concentrated Wash Solution.

Dilute 30 mL of concentrated *Wash Solution* with 1170 mL deionized water to a final volume of 1200 mL.

The diluted Wash Solution is stable for 2 weeks at room temperature.

Enzyme Conjugate

Dilute *Enzyme Conjugate* concentrate 1:21 in *Conjugate Diluent*.

Stability of the prepared Enzyme-Conjugate: Use within 24 hours.

Example:

If the whole plate is used, dilute 300 µL *Enzyme Conjugate* (21x conc.) with 6 mL *Conjugate Diluent* to a total volume of 6.3 mL.

If the whole plate is not used at once prepare the required quantity of enzyme conjugate by mixing 25 µL of *Enzyme Conjugate* 21X conc. with 0.5 mL of *Conjugate Diluent* per strip (see table below):

| No. of strips | <i>Enzyme Conjugate</i> 21X conc. (µL) | <i>Conjugate Diluent</i> (mL) |
|---------------|--|-------------------------------|
| 1 | 25 | 0.5 |
| 2 | 50 | 1.0 |
| 3 | 75 | 1.5 |
| 4 | 100 | 2.0 |
| 5 | 125 | 2.5 |
| 6 | 150 | 3.0 |
| 7 | 175 | 3.5 |
| 8 | 200 | 4.0 |
| 9 | 225 | 4.5 |
| 10 | 250 | 5.0 |
| 11 | 275 | 5.5 |
| 12 | 300 | 6.0 |

Chemiluminescence Substrate Solution

Mix **1 part** of the chemiluminescence **Reagent A** with **1 part** of **Reagent B** and dilute this mixture 1:2 with **Reagent C**. This gives the ready to use substrate solution.

The prepared substrate solution is stable for one hour. Prepare fresh before use.

If the whole plate is to be used prepare the substrate solution as follows:

Add 1.5 mL of each *Reagent A* and *Reagent B* into 3 mL *Reagent C*.

4.5 Disposal of the Kit

The disposal of the kit must be made according to the national regulations. Special information for this product is given in the Material Safety Data Sheets (see chapter 13).

4.6 Damaged Test Kits

In case of any severe damage to the test kit or components, DRG has to be informed in writing, at the latest, one week after receiving the kit. Severely damaged single components should not be used for a test run. They have to be stored until a final solution has been found. After this, they should be disposed according to the official regulations.

5 SPECIMEN COLLECTION AND PREPARATION

Serum or Heparin plasma can be used in this assay.

Citrate plasma results in decreased, EDTA in strongly increased values.

Do not use haemolytic, icteric or lipaemic specimens.

Please note: Samples containing sodium azide should not be used in the assay.

5.1 Specimen Collection

Serum:

Collect blood by venipuncture (e.g. Sarstedt Monovette # 02.1388.001), allow to clot, and separate serum by centrifugation at room temperature. Do not centrifuge before complete clotting has occurred. Patients receiving anticoagulant therapy may require increased clotting time.

Plasma:

Whole blood should be collected into centrifuge tubes containing anti coagulant and centrifuged immediately after collection.

(E.g. for Heparin plasma Sarstedt Monovette – orange cap - # 02.165.001)

5.2 Specimen Storage and Preparation

Specimens should be capped and may be stored for up to 2 days at 2°C to 8°C prior to assaying.

Specimens held for a longer time should be frozen only once at -20°C prior to assay. Thawed samples should be inverted several times prior to testing.

5.3 Specimen Dilution

If in an initial assay, a specimen is found to contain more than the highest standard, the specimens can be diluted with *Sample Diluent* and reassayed as described in Assay Procedure.

For the calculation of the concentrations this dilution factor has to be taken into account.

Example:

- a) Dilution 1:10: 10 µL Serum + 90 µL *Sample Diluent* (mix thoroughly)
- b) Dilution 1:100: 10 µL dilution a) 1:10 + 90 µL *Sample Diluent* (mix thoroughly).

6 ASSAY PROCEDURE

6.1 General Remarks

- All reagents and specimens must be allowed to come to room temperature before use. All reagents must be mixed without foaming.
- Once the test has been started, all steps should be completed without interruption.
- Use new disposal plastic pipette tips for each standard, control or sample in order to avoid cross contamination.
- Light intensity is a function of the incubation time and temperature. Before starting the assay, it is recommended that all reagents are ready, caps removed, all needed wells secured in holder, etc. This will ensure equal elapsed time for each pipetting step without interruption.
- As a general rule the enzymatic reaction is linearly proportional to time and temperature.

6.2 Test Procedure

Each run must include a standard curve.

1. Secure the desired number of Microtiter wells in the frame holder.
2. Dispense **50 µL** of each **Standard, Control** and **samples** with new disposable tips into appropriate wells.
3. Dispense **50 µL** freshly diluted Enzyme Conjugate (see "Reagent Preparation") into each well. Thoroughly mix for 10 seconds. It is important to have a complete mixing in this step.
4. Incubate for **30 minutes** at room temperature (without covering the plate).
5. Briskly shake out the contents of the wells.
Rinse the wells 3 times with diluted *Wash Solution* (300 µL per well). Strike the wells sharply on absorbent paper to remove residual droplets.
Important note:
The sensitivity and precision of this assay is markedly influenced by the correct performance of the washing procedure!
6. Add **50 µL** of the freshly prepared Substrate Solution to each well. (See "Reagent Preparation".)
7. Incubate for **10 minutes** at room temperature.
8. Read the RLU with a microtiter plate luminometer **within 20 minutes** after incubation time of substrate.

6.3 Calculation of Results

1. Calculate the average Relativ Light Units (RLU) values for each set of standards, controls and patient samples.
2. Using semi-logarithmic graph paper, construct a standard curve by plotting the mean RLU obtained from each standard against its concentration with RLU value on the vertical(Y) axis and concentration on the horizontal (X) axis.
3. Using the mean RLU value for each sample determine the corresponding concentration from the standard curve.
4. Automated method: The results in the IFU have been calculated automatically using a 4 PL (4 Parameter Logistics) curve fit. 4 Parameter Logistics is the preferred method. Other data reduction functions may give slightly different results.
5. The concentration of the samples can be read directly from this standard curve. Samples with concentrations higher than that of the highest standard have to be further diluted or reported as > 50 ng/mL. For the calculation of the concentrations this dilution factor has to be taken into account.

6.3.1 Example of Typical Standard Curve

The following data** is for demonstration only and **cannot** be used in place of data generations at the time of assay.

| Standard [ng/mL] | RLU ($\times 10^3$) | RLU/RLU _{max} [%] |
|------------------|-----------------------|----------------------------|
| Standard 0 (0) | 8.7 | 0.3 |
| Standard 1 (3) | 447 | 16.9 |
| Standard 2 (10) | 1186 | 44.9 |
| Standard 3 (25) | 2012 | 76.1 |
| Standard 4 (50) | 2642 | 100.0 |

** It is recommended to use the RLU/RLU_{max} values for comparative purposes since luminometers vary considerably between manufacturers. Results from different luminometers will show different RLU values, however, the RLU/RLU_{max} values remain consistent.

7 EXPECTED NORMAL VALUES

It is strongly recommended that each laboratory should determine its own normal and abnormal values.

| Population | Valid N | 5% Percentile | 95% Percentile |
|---------------|---------|---------------|----------------|
| Men and women | 71 | 0.37 ng/mL | 2.82 ng/mL |

The results alone should not be the only reason for any therapeutic consequences. The results should be correlated to other clinical observations and diagnostic tests.

8 QUALITY CONTROL

Good laboratory practice requires that controls be run with each calibration curve. A statistically significant number of controls should be assayed to establish mean values and acceptable ranges to assure proper performance.

It is recommended to use control samples according to state and federal regulations. The use of control samples is advised to assure the day to day validity of results. Use controls at both normal and pathological levels.

The controls and the corresponding results of the QC-Laboratory are stated in the QC certificate added to the kit. The values and ranges stated on the QC sheet always refer to the current kit lot and should be used for direct comparison of the results.

It is also recommended to make use of national or international Quality Assessment programs in order to ensure the accuracy of the results.

Employ appropriate statistical methods for analysing control values and trends. If the results of the assay do not fit to the established acceptable ranges of control materials patient results should be considered invalid. In this case, please check the following technical areas: Pipetting and timing devices; photometer, expiration dates of reagents, storage and incubation conditions, aspiration and washing methods.

After checking the above mentioned items without finding any error contact your distributor or DRG directly.

9 PERFORMANCE CHARACTERISTICS

9.1 Assay Dynamic Range

The range of the assay is between 0.03 – 50 ng/mL.

9.2 Specificity of Antibodies (Cross Reactivity)

Sera of healthy individuals did not react with the DRG CYFRA 21-1 CLIA

9.3 Sensitivity

The analytical sensitivity of the DRG CLIA was calculated by adding 2 standard deviations to the mean of 20 replicate analyses of *Standard 0* and was found to be 0.03 ng/mL.

9.4 Reproducibility

9.4.1 Intra Assay

The within assay variability is shown below:

| Samples | 1 | 2 | 3 |
|--------------|------|------|------|
| Mean [ng/mL] | 2.43 | 5.10 | 9.48 |
| CV (%) | 2.68 | 1.38 | 2.70 |
| n = | 20 | 20 | 20 |

9.4.2 Inter Assay

The between assay variability is shown below:

| Samples | 1 | 2 | 3 |
|--------------|------|------|-------|
| Mean [ng/mL] | 3.66 | 7.85 | 12.25 |
| CV (%) | 4.24 | 2.65 | 4.25 |
| n = | 18 | 18 | 18 |

9.5 Recovery

Recovery was determined by adding increasing amounts of the analyt to three different sera containing different amounts of endogenous analyt. Each sample (non-spiked and spiked) was assayed and analyt concentrations of the samples were calculated from the standard curve. The percentage recoveries were determined by comparing expected and measured values of the samples.

| | | Sample 1 | Sample 2 | Sample 3 |
|-----------------------|------|----------|----------|----------|
| Concentration [ng/mL] | | 15.45 | 10.38 | 11.69 |
| Average Recovery | | 96.8 | 89.3 | 91.0 |
| Range of Recovery [%] | from | 86.4 | 85.4 | 87.5 |
| | to | 106.2 | 94.8 | 97.3 |

9.6 Linearity

| | | Sample 1 | Sample 2 | Sample 3 |
|-----------------------|------|----------|----------|----------|
| Concentration [ng/mL] | | 14.45 | 9.65 | 10.46 |
| Average Recovery | | 80.7 | 70.1 | 74.6 |
| Range of Recovery [%] | from | 42.1 | 28.2 | 27.5 |
| | to | 107.8 | 94.7 | 97.9 |

10 LIMITATIONS OF USE

Reliable and reproducible results will be obtained when the assay procedure is performed with a complete understanding of the package insert instruction and with adherence to good laboratory practice. Any improper handling of samples or modification of this test might influence the results.

10.1 Interfering Substances

Haemoglobin (up to 4 mg/mL), Bilirubin (up to 0.5 mg/mL) and Triglyceride (up to 7.5 mg/mL) have no influence on the assay results.

The assay contains reagents to minimize interference of HAMA and heterophilic antibodies. However, extremely high titers of HAMA or heterophilic antibodies may interfere with the test results.

10.2 Drug Interferences

Until today no substances (drugs) are known to us, which have an influence to the measurement of CYFRA 21-1 in a sample.

10.3 High-Dose-Hook Effect

No hook effect was observed in this test up to 400 ng/mL of CYFRA 21-1.

11 LEGAL ASPECTS

11.1 Reliability of Results

The test must be performed exactly as per the manufacturer's instructions for use. Moreover the user must strictly adhere to the rules of GLP (Good Laboratory Practice) or other applicable national standards and/or laws. This is especially relevant for the use of control reagents. It is important to always include, within the test procedure, a sufficient number of controls for validating the accuracy and precision of the test.

The test results are valid only if all controls are within the specified ranges and if all other test parameters are also within the given assay specifications. In case of any doubt or concern please contact DRG.

11.2 Therapeutic Consequences

Therapeutic consequences should never be based on laboratory results alone even if all test results are in agreement with the items as stated under point 11.1. Any laboratory result is only a part of the total clinical picture of a patient.

Only in cases where the laboratory results are in acceptable agreement with the overall clinical picture of the patient should therapeutic consequences be derived.

The test result itself should never be the sole determinant for deriving any therapeutic consequences.

11.3 Liability

Any modification of the test kit and/or exchange or mixture of any components of different lots from one test kit to another could negatively affect the intended results and validity of the overall test. Such modification and/or exchanges invalidate any claim for replacement.

Claims submitted due to customer misinterpretation of laboratory results subject to point 11.2. are also invalid. Regardless, in the event of any claim, the manufacturer's liability is not to exceed the value of the test kit. Any damage caused to the test kit during transportation is not subject to the liability of the manufacturer.

12 REFERENCES

1. Petra Stieber CYFRA 21-1 (Cytokeratin-19-Fragment) in Lothar Thomas, Labor and Diagnose, TH Books, Frankfurt, Germany.
2. J-L Pujol, O Molinier, W Ebert, J-P Daures, F Barlesi, G Bucceri, M Paesmans, E Quoix, D Moro-Sibilot, M Szturmowicz, J-M Brechot, T Muley and J Grenier (2004) British Journal of Cancer 90(11):2097-2105.

1 EINLEITUNG

Der **DRG CYFRA 21-1 CLIA** wird zur quantitativen Bestimmung von CYFRA 21-1 in Serum und Heparinplasma eingesetzt.

Nur für In-vitro Diagnostik.

2 TESTPRINZIP

Der DRG CYFRA 21-1 CLIA ist ein **Chemilumineszenz-Immunoassay (CLIA)**, der auf der Sandwichtechnik basiert.

Die Wells der Mikrotiterplatten sind mit einem monoklonalen Antikörper beschichtet, der gegen eine definierte Antikörper-Bindungsstelle des CYFRA 21-1-Moleküls gerichtet ist.

Die Proben werden in die beschichteten Wells gegeben und mit einem Enzymkonjugat inkubiert. Das Konjugat enthält einen anti-CYFRA 21-1-Antikörper, der mit Meerrettichperoxidase konjugiert ist. Es wird ein Sandwichkomplex gebildet.

Das nicht gebundene Konjugat wird durch Waschen der Wells entfernt. Nach Zugabe der Substratlösung ist die Intensität des emittierten Lichtes proportional der CYFRA 21-1-Konzentration in der Probe.

3 VORSICHTSMAßNAHMEN

- Dieser Kit ist nur zum in vitro diagnostischen Gebrauch geeignet.
- Alle Bestandteile dieses Testkits, die humanes Serum oder Plasma enthalten, wurden mit FDA-geprüften Methoden auf HIV I/II, HbsAg und HCV getestet und als negativ bestätigt. Jedoch sollten alle Bestandteile im Umgang und bei der Entsorgung wie mögliche Gefahrenstoffe betrachtet werden.
- Nicht mit dem Mund pipettieren und den Kontakt von Kitbestandteilen und Proben mit Haut und Schleimhäuten vermeiden.
- In den Bereichen, in denen Proben oder Kitbestandteile verwendet werden, nicht rauchen, essen oder Kosmetika verwenden.
- Beim Umgang mit Proben oder Reagenzien Einweg-Latexhandschuhe tragen. Die Verunreinigung von Reagenzien oder Proben mit Mikroben kann zu falschen Ergebnissen führen.
- Der Gebrauch sollte gemäß der Vorschriften einer entsprechenden nationalen Gefahrenstoff-Sicherheitsrichtlinie erfolgen.
- Reagenzien nicht nach dem auf dem Kit-Etikett angegebenen Verfallsdatum verwenden.
- Alle im Kit-Protokoll angegebenen Mengen müssen genau eingehalten werden. Optimale Ergebnisse können nur durch Verwendung kalibrierter Pipetten und Mikrotiterplatten-Lesegeräte erreicht werden.
- Die Lumineszenz-Substrat-Reagenzien (*Reagent A* und *Reagent B*) sind lichtempfindlich und müssen vor Sonnenlicht geschützt im dunklen Originalbehälter aufbewahrt werden.
- Komponenten von Kits mit unterschiedlichen Lotnummern nicht untereinander vertauschen. Es wird empfohlen, keine Wells von verschiedenen Platten zu verwenden, auch nicht, wenn es sich um das gleiche Lot handelt. Die Kits können unter anderen Bedingungen gelagert oder versendet worden sein, so dass die Bindungscharakteristik der Platten leicht unterschiedlich ausfällt.
- Chemikalien und zubereitete oder bereits benutzte Reagenzien müssen gemäß den nationalen Gefahrenstoffvorschriften wie gefährlicher Abfall behandelt werden.
- Informationen zu im Kit enthaltenen gefährlichen Substanzen entnehmen Sie bitte dem Material Sicherheitsdatenblatt.
- Material Sicherheitsdatenblätter für dieses Produkt sind auf Anfrage direkt von der Firma DRG Instruments GmbH erhältlich. Die Material Sicherheitsdatenblätter entsprechen den Verordnungen der EU-Richtlinie 91/155 EC.

4 BESTANDTEILE DES KITS

4.1 Kitinhalt

1. **Microtiterwells**, 96 Wells, 12x8 Wells;
mit anti- CYFRA 21-1 -Antikörper (monoklonal) beschichtet.
 2. **Standard (Standard 0-4)**, 5 Fläschchen (lyophilisiert), je 1 mL
Konzentrationen: 0 – 3 – 10 – 25 – 50 ng/mL
siehe „Vorbereitung der Reagenzien“
* Enthält 0,03% Proclin 300, 0.015% BND und 0.010% MIT als Konservierungsmittel.
 3. **Control** (Kontrolle), 1 Fläschchen (lyophilisiert), 1 mL,
siehe „Vorbereitung der Reagenzien“,
Kontrollwerte und –bereiche entnehmen Sie bitte dem Fläschchenetikett oder dem QC-Datenblatt.
* Enthält 0,03% Proclin 300, 0.015% BND und 0.010% MIT als Konservierungsmittel.
 4. **Sample Diluent** (Probenverdünnungsmedium), 1 Fläschchen, 3 mL, gebrauchsfertig.
* Enthält 0,03% Proclin 300, 0.015% BND und 0.010% MIT als Konservierungsmittel.
 5. **Enzyme Conjugate** (Enzymkonjugat)**21X konzentriert**, 1 Fläschchen, 0,5 mL,
Anti-CYFRA 21-1-Antikörper, mit Meerrettichperoxidase konjugiert,
siehe „Vorbereitung der Reagenzien“
* Enthält 0,03% Proclin 300, 0.015% BND und 0.010% MIT als Konservierungsmittel.
 6. **Conjugate Diluent** (Konjugatverdünnungsmedium), 1 Fläschchen, 7 mL, gebrauchsfertig.
* Enthält 0,03% Proclin 300, 0.015% BND und 0.010% MIT als Konservierungsmittel.
 7. **Substrate Solution** (Substratlösung),
Reagent A, 1 Fläschchen, 2,0 mL, *Achtung: Lichtempfindlich!*
Reagent B, 1 Fläschchen, 2,0 mL, *Achtung: Lichtempfindlich!*
Reagent C, 1 Fläschchen, 5,0 mL
Siehe „Vorbereitung der Reagenzien“.
 8. **Wash Solution** (Waschlösung), **40X** konzentriert, 1 Fläschchen, 30 mL;
Siehe „Vorbereitung der Reagenzien“.
- * BND = 5-bromo-5-nitro-1,3-dioxane
MIT = 2-methyl-2H-isothiazol-3-one

Anmerkung: Zusätzliches *Sample Diluent* zur Probenverdünnung ist auf Anfrage erhältlich.

4.2 Nicht im Kit enthaltene aber erforderliche Geräte und Materialien

- Mikrotiterplatten-Luminometer
- Kalibrierte variable Präzisions-Mikropipette
- Saugfähiges Papier
- Aqua dest.

4.3 Lagerung und Haltbarkeit des Kits

Die ungeöffneten Reagenzien behalten bei Lagerung um 2°C bis 8°C ihre Reaktivität bis zum Verfallsdatum m.

Nach dem Verfallsdatum die Reagenzien nicht mehr verwenden.

Nach dem Öffnen sollten alle Reagenzien bei 2°C bis 8°C gelagert werden.

Die Mikrotiterwells sollten bei 2°C bis 8°C gelagert werden. Der einmal geöffnete Folienbeutel sollte stets sehr sorgfältig wieder verschlossen werden.

Achtung: Bei 2°C bis 8°C sind die rekonstituierten Standards und Kontrollen mindestens 4 Wochen haltbar.

Für eine längere Aufbewahrung bei -20°C einfrieren.

4.4 Vorbereitung der Reagenzien

Alle Reagenzien sowie die benötigte Anzahl von Wells sollen vor dem Gebrauch auf Raumtemperatur gebracht werden.

Standards

Rekonstituieren Sie den lyophilisierten Inhalt der Standardfläschchen mit 1 mL destilliertem Wasser.

Achtung: Bei 2°C bis 8°C sind die rekonstituierten Standards mindestens 4 Wochen haltbar. Für eine längere Aufbewahrung bei -20°C einfrieren.

Control

Rekonstituieren Sie den lyophilisierten Inhalt des Fläschchens mit 1 mL destilliertem Wasser und lassen Sie das Fläschchen mindestens 10 Minuten ruhen. Vor Gebrauch die Kontrolle mehrmals vorsichtig schütteln.

Achtung: Bei 2°C bis 8°C ist die rekonstituierte Kontrolle mindestens 4 Wochen haltbar. Für eine längere Aufbewahrung bei -20°C einfrieren.

Wash Solution

Die 40-fach konzentrierte Wash Solution (30 mL) mit 1170 mL destilliertem Wasser auf ein Gesamtvolumen von 1200 mL verdünnen. Die verdünnte Waschlösung ist bei Raumtemperatur für 2 Wochen stabil.

Enzyme Conjugate

Verdünnen Sie das Enzymkonjugat-Konzentrat 1:21 mit Conjugate Diluent.

Stabilität der Enzymkonjugat-Arbeitslösung: Innerhalb von 24 Stunden verwenden.

Beispiel:

Wird die ganze Platte verwendet, verdünnen Sie 300 µL Enzyme Conjugate 21X mit 6 mL Conjugate Diluent (Gesamtvolumen 6,3 mL).

Wird nur ein Teil der Platte benötigt, setzen Sie nur das erforderliche Volumen des Konjugates an. Verdünnen Sie dafür pro Streifen 25 µL Enzyme Conjugate-Konzentrat mit 0,5 mL Conjugate Diluent (siehe Tabelle):

| Anzahl der Streifen | Enzyme Conjugate 21X konz. (µL) | Conjugate Diluent (mL) |
|---------------------|---------------------------------|------------------------|
| 1 | 25 | 0,5 |
| 2 | 50 | 1,0 |
| 3 | 75 | 1,5 |
| 4 | 100 | 2,0 |
| 5 | 125 | 2,5 |
| 6 | 150 | 3,0 |
| 7 | 175 | 3,5 |
| 8 | 200 | 4,0 |
| 9 | 225 | 4,5 |
| 10 | 250 | 5,0 |
| 11 | 275 | 5,5 |
| 12 | 300 | 6,0 |

Chemiluminescence Substrate Solution

Mischen Sie 1 Teil Reagent A mit 1 Teil Reagent B. Verdünnen Sie diese Mischung 1:2 mit Reagent C. Dadurch erhalten Sie die gebrauchsfertige Substratlösung.

Diese angesetzte Substratlösung ist für 1 Stunde stabil. Die Lösung muss immer frisch angesetzt werden.

Wird die ganze Platte verwendet, setzen sie die Substratlösung wie folgt an:

1,5 ml Reagent A + 1,5 ml Reagent B in 3,0 mL Reagent C geben.

4.5 Entsorgung des Kits

Die Entsorgung des Kits muss gemäß den nationalen gesetzlichen Vorschriften erfolgen. Spezielle Informationen für dieses Produkt finden Sie im Material Sicherheitsdatenblatt, Kapitel 13.

4.6 Beschädigte Testkits

Im Falle einer starken Beschädigung des Testkits oder der Komponenten muss die Firma DRG in schriftlicher Form spätestens eine Woche nach Erhalt des Kits informiert werden. Stark beschädigte Einzelkomponenten sollten nicht für den Testlauf verwendet werden. Sie müssen gelagert werden bis eine endgültige Lösung gefunden wurde. Danach sollten Sie gemäß den offiziellen Richtlinien entsorgt werden.

5 PROBENVORBEREITUNG

Serum oder Heparinplasma kann in diesem Test als Probenmaterial eingesetzt werden. Zitratplasma führt zu erniedrigten und EDTA-Plasma zu deutlich erhöhten Werten.

Lipämische, ikterische und/oder stark hämolysierte Proben sollten nicht verwendet werden.

Achtung: Proben, die Natriumazid enthalten, sollten nicht verwendet werden.

5.1 Probenentnahme

Serum:

Blut durch Venenpunktion entnehmen (z.B. mit Sarstedt Monovette # 02.1388.001), gerinnen lassen und das Serum durch Zentrifugation bei Raumtemperatur abtrennen. Vor der Zentrifugation muss die Gerinnung vollständig abgeschlossen sein. Bei Patienten, die Antikoagulantien erhalten, kann die Gerinnungszeit länger dauern.

Plasma:

Die Blutentnahme erfolgt mit Röhrchen, die ein Antikoagulantien enthalten. Das Plasma wird als Überstand nach einer Zentrifugation gewonnen.

(z.B. für Heparinplasma Sarstedt Monovette – oranger Deckel - # 02.165.001)

5.2 Probenaufbewahrung

Proben sollten stets gut verschlossen sein und können vor Testbeginn bis zu 2 Tage bei 2°C bis 8°C gelagert werden.

Für eine längere Aufbewahrung sollten die Proben eingefroren bei -20°C bis zum Testbeginn gelagert werden. Nur einmal einfrieren. Aufgetaute Proben sollten vor Testbeginn vorsichtig durchmischt werden, ohne Schaumbildung.

5.3 Probenverdünnung

Wenn in einem ersten Testdurchlauf bei einer Probe eine Konzentration höher als der höchste Standard gefunden wird, kann diese Probe mit *Sample Diluent* weiter verdünnt und nochmals bestimmt werden. Die Verdünnung muss jedoch bei der Berechnung der Konzentration beachtet werden.

Beispiel:

- a) Verdünnung 1:10: 10 µL Serum + 90 µL *Sample Diluent* gründlich mischen)
- b) Verdünnung 1:100: 10 µL Verdünnung a) 1:10 + 90 µL *Sample Diluent* (gründlich mischen).

6 TESTDURCHFÜHRUNG

6.1 Allgemeine Hinweise

- Alle Reagenzien und Proben müssen vor Gebrauch auf Raumtemperatur gebracht und gut durchmischt werden. Dabei sollte Schaumbildung vermieden werden.
- Wenn die Testdurchführung einmal begonnen wurde, muss sie ohne Unterbrechung zu Ende geführt werden.
- Für jeden Standard, jede Kontrolle oder Probe eine neue Plastikspitze verwenden, um Verschleppungen zu vermeiden.
- Die Lichtintensität ist abhängig von Inkubationszeit und Temperatur. Deshalb ist es notwendig, vor Beginn der Testdurchführung alle Reagenzien in einen arbeitsbereiten Zustand zu bringen, die Deckel der Fläschchen zu öffnen, alle benötigten Wells in den Halter zu setzen. Nur eine solche Vorbereitung garantiert gleiche Zeiten für jeden Pipettiervorgang ohne Pausen.
- Als generelle Regel gilt, dass die enzymatische Reaktion linear proportional zu Zeit und Temperatur ist.

6.2 Testdurchführung

Jeder Lauf muss eine Standardkurve beinhalten.

1. Die benötigte Anzahl Wells in der Halterung befestigen.
2. **Je 50 µL Standard, Control** und Proben mit neuen Plastikspitzen in die entsprechenden Wells geben.
3. **50 µL frisch angesetztes Enzymkonjugat** (siehe „Vorbereitung der Reagenzien“) in jedes Well geben. Für 10 Sekunden gut schütteln. Es ist sehr wichtig, in diesem Schritt eine komplette Durchmischung zu erreichen.
4. **30 Minuten** bei Raumtemperatur inkubieren (Platte nicht abdecken).
5. Den Inhalt der Wells kräftig ausschütteln. Wells 3mal mit verdünnter *Wash Solution* (300 µL) waschen. Verbleibende Flüssigkeit durch Ausklopfen der Wells auf saugfähigem Papier entfernen.
Achtung: Die Sensitivität und Präzision dieses Assays wird erheblich beeinflusst von der korrekten Durchführung des Waschschrilles!
6. **50 µL frisch zubereitete** Substratlösung in jedes Well geben. (Siehe „Vorbereitung der Reagenzien“.)
7. **10 Minuten** bei Raumtemperatur inkubieren.
8. Messen der RLU mit einem Mikrotiterplatten-Luminometer **innerhalb von 20 Minuten** nach Ende der Substrat-Inkubation

6.3 Ergebnisermittlung

1. Die durchschnittlichen Werte der „Relativ Light Units“ (RLU) für jedes Set von Standards, Controls und Patientenproben bestimmen.
2. Eine Standardkurve ermitteln durch Auftragen der mittleren RLU jedes Standards gegen die Konzentration, wobei der RLU-Wert auf der vertikalen (Y) Achse und die Konzentration auf der horizontalen (X) Achse eingetragen wird.
3. Unter Verwendung der mittleren RLU wird für jede Probe die entsprechende Konzentration aus der Standardkurve ermittelt.
4. Automatische Methode: Die in der Arbeitsanleitung ermittelten Werte wurden automatisch mit Hilfe der 4 Parameter Gleichung (4 PL, 4 Parameter Logistics, 4 Parameter Rodbard) bestimmt. Andere Auswertungsfunktionen können leicht abweichende Werte ergeben.
5. Die Konzentration der Proben kann direkt von der Standardkurve abgelesen werden. Proben, die eine höhere Konzentration als die des höchsten Standards enthalten, müssen verdünnt werden. Dieser Verdünnungsfaktor muss bei der Berechnung der Konzentration beachtet werden.

6.3.1 Beispiel für eine Standardkurve

Nachfolgend wird ein typisches Beispiel für eine Standardkurve** mit dem DRG CLIA gezeigt. Diese Werte sollten **nicht** zur Berechnung von Patientendaten verwendet werden.

| Standard [ng/mL] | RLU ($\times 10^3$) | RLU/RLU _{max} [%] |
|------------------|-----------------------|----------------------------|
| Standard 0 (0) | 8,7 | 0,3 |
| Standard 1 (3) | 447 | 16,9 |
| Standard 2 (10) | 1186 | 44,9 |
| Standard 3 (25) | 2012 | 76,1 |
| Standard 4 (50) | 2642 | 100,0 |

** Um vergleichbare Werte zu erhalten, wird empfohlen, das Verhältnis von RLU/RLU_{max} zu verwenden, da die Messungen zwischen Luminometern verschiedener Hersteller variieren. Verschiedene Luminometer ergeben unterschiedliche RLU-Werte, das Verhältnis von RLU/RLU_{max} dagegen bleibt konstant.

7 ERWARTETE WERTE

Es wird empfohlen, dass jedes Labor seine eigenen normalen und abnormalen Werte ermittelt.

| Population | N | 5% Perzentile | 95% Perzentile |
|-------------------|----|---------------|----------------|
| Männer und Frauen | 71 | 0,37 ng/mL | 2,82 ng/mL |

8 QUALITÄTS-KONTROLLE

Es wird empfohlen, die Kontrollproben gemäß den nationalen gesetzlichen Bestimmungen einzusetzen. Durch die Verwendung von Kontrollproben wird eine Tag-zu-Tag Überprüfung der Ergebnisse erzielt. Es sollten Kontrollen sowohl mit normalem als auch pathologischem Level eingesetzt werden.

Die Kontrollen mit den entsprechenden Ergebnissen des QC-Labors sind im QC-Zertifikat, das dem Kit beiliegt, aufgeführt. Die im QC-Blatt angegebenen Werte und Bereiche beziehen sich stets auf die aktuelle Kitcharge und sollten zum direkten Vergleich der Ergebnisse verwendet werden.

Es wird ebenfalls empfohlen, an nationalen oder internationalen Qualitätssicherungs-Programmen teilzunehmen, um die Genauigkeit der Ergebnisse zu sichern.

Es sollten geeignete statistische Methoden zur Analyse von Kontroll-Werten und Trends angewendet werden. Wenn die Ergebnisse des Assays nicht mit den angegebenen Akzeptanzbereichen des Kontrollmaterials übereinstimmen, sollten die Patientenergebnisse als ungültig eingestuft werden.

In diesem Fall überprüfen Sie bitte die folgenden Bereiche: Pipetten und Zeitnehmer, Photometer, Verfallsdatum der Reagenzien, Lagerungs- und Inkubationsbedingungen, Absaug- und Waschmethode. Sollten Sie nach Überprüfung der vorgenannten Bereiche keinen Fehler erkannt haben, setzen Sie sich bitte mit Ihrem Lieferanten oder direkt mit der Firma DRG in Verbindung.

9 ASSAY CHARACTERISTIKA

9.1 Messbereich

Der Messbereich des Testes liegt zwischen 0,03 – 50 ng/mL.

9.2 Spezifität der Antikörper (Kreuzreaktivität)

Seren von gesunden Individuen reagierten nicht mit dem DRG CYFRA 21-1-ELISA

9.3 Analytische Sensitivität

Die analytische Sensitivität, definiert als Mittelwert plus der zweifachen Standardabweichung des *Standards 0* (n = 20), beträgt 0,03 ng/mL.

Die Daten zu:

9.4 Präzision

9.5 Wiederfindung

9.6 Linearität

entnehmen Sie bitte der ausführlichen englischen Arbeitsanleitung.

10 GRENZEN DES TESTS

Jede unsachgemäße Behandlung von Proben oder Modifikationen dieses Tests können die Ergebnisse beeinflussen.

10.1 Interferenzen

Hämoglobin (bis zu 4 mg/mL), Bilirubin (bis zu 0,5 mg/mL) und Triglyceride (bis zu 7,5 mg/mL) haben keinen Einfluss auf das Testergebnis.

Der Test enthält Reagenzien, um Interferenzen mit HAMA oder heterophilen Antikörpern zu minimieren.

Trotzdem ist es möglich, dass ein sehr hoher Titer von HAMA oder heterophilen Antikörpern das Testergebnis beeinflusst.

10.2 Beeinflussung durch Medikamente

Uns sind bislang keine Stoffe (Medikamente) bekannt geworden, deren Einnahme die Messung des CYFRA 21-1-Gehaltes der Probe beeinflussen würde.

10.3 High-Dose-Hook Effekt

Ein Hook Effekt tritt bei Proben mit bis zu 400 ng/mL CYFRA 21-1 nicht auf.

11 RECHTLICHE GRUNDLAGEN

11.1 Zuverlässigkeit der Ergebnisse

Der Test muss exakt gemäß der Testanleitung des Herstellers abgearbeitet werden. Darüber hinaus muss der Benutzer sich strikt an die Regeln der GLP (Good Laboratory Practice) oder andere eventuell anzuwendende Regeln oder nationale gesetzliche Vorgaben halten. Dies betrifft besonders den Gebrauch der Kontrollreagenzien. Es ist sehr wichtig, bei der Testdurchführung stets eine ausreichende Anzahl Kontrollen zur Überprüfung der Genauigkeit und Präzision mitlaufen zu lassen.

Die Testergebnisse sind nur gültig, wenn alle Kontrollen in den vorgegebenen Bereichen liegen, und wenn alle anderen Testparameter die vorgegebenen Spezifikationen für diesen Assay erfüllen. Wenn Sie bezüglich eines Ergebnisses Zweifel oder Bedenken haben, setzen Sie sich bitte mit der Firma DRG in Verbindung.

11.2 Therapeutische Konsequenzen

Therapeutische Konsequenzen sollten keinesfalls nur aufgrund von Laborergebnissen erfolgen, selbst dann nicht, wenn alle Testergebnisse mit den in 11.1. genannten Voraussetzungen übereinstimmen. Jedes Laborergebnis ist nur ein Teil des klinischen Gesamtbildes eines Patienten.

Nur in Fällen, in denen die Laborergebnisse in akzeptabler Übereinstimmung mit dem allgemeinen klinischen Bild des Patienten stehen, sollten therapeutische Konsequenzen eingeleitet werden.

Das Testergebnis allein sollte niemals als alleinige Grundlage für die Einleitung therapeutischer Konsequenzen dienen.

11.3 Haftung











Jegliche Veränderungen des Testkits und/oder Austausch oder Vermischung von Komponenten unterschiedlicher Chargen von einem Testkit zu einem anderen, können die gewünschten Ergebnisse und die Gültigkeit des gesamten Tests negativ beeinflussen. Solche Veränderungen und/oder Austausch haben den Ausschluss jeglicher Ersatzansprüche zur Folge.






Reklamationen, die aufgrund von Falschinterpretation von Laborergebnissen durch den Kunden gemäß Punkt 11.2. erfolgen, sind ebenfalls abzuweisen. Im Falle jeglicher Reklamation ist die Haftung des Herstellers maximal auf den Wert des Testkits beschränkt. Jegliche Schäden, die während des Transports am Kit entstanden sind, unterliegen nicht der Haftung des Herstellers.

12 REFERENZEN

Angaben zu den Referenzen entnehmen Sie bitte der ausführlichen englischen Arbeitsanleitung.

SYMBOLS USED WITH DRG CLIAS

| Symbol | English | Deutsch | Français | Español | Italiano |
|---|------------------------------------|------------------------------|--|---------------------------------------|-------------------------------------|
|  | European Conformity | CE-Konformitätskennzeichnung | Conforme aux normes européennes | Conformidad europea | Conformità europea |
|  | Consult instructions for use | Gebrauchsanweisung beachten | Consulter les instructions d'utilisation | Consulte las Instrucciones | Consultare le istruzioni per l'uso |
|  | In vitro diagnostic device | In-vitro-Diagnostikum | Usage Diagnostic in vitro | Para uso Diagnóstico in vitro | Per uso Diagnostica in vitro |
|  | For research use only | Nur für Forschungszwecke | Seulement dans le cadre de recherches | Sólo para uso en investigación | Solo a scopo di ricerca |
|  | Catalogue number | Katalog-Nr. | Référence | Número de catálogo | No. di Cat. |
|  | Lot. No. / Batch code | Chargen-Nr. | No. de lot | Número de lote | Lotto no |
|  | Contains sufficient for <n> tests/ | Ausreichend für "n" Ansätze | Contenu suffisant pour "n" tests | Contenido suficiente para <n> ensayos | Contenuto sufficiente per "n" saggi |
|  | Storage Temperature | Lagerungstemperatur | Température de conservation | Temperatura de conservación | Temperatura di conservazione |
|  | Expiration Date | Mindesthaltbarkeitsdatum | Date limite d'utilisation | Fecha de caducidad | Data di scadenza |
|  | Legal Manufacturer | Hersteller | Fabricant | Fabricante | Fabbricante |
| <i>Distributed by</i> | Distributor | Vertreiber | Distributeur | Distribuidor | Distributore |
| <i>Content</i> | Content | Inhalt | Contenu | Contenido | Contenuto |
| <i>Volume/No.</i> | Volume / No. | Volumen/Anzahl | Volume/Numéro | Volumen/Número | Volume/Quantità |
| <i>Microtiterwells</i> | Microtiterwells | Mikrotiterwells | Plaques de micro-titration | Placas multipocillo | Micropozzetti |
| <i>Antiserum</i> | Antiserum | Antiserum | Antisérum | Antisuero | Antisiero |
| <i>Enzyme Conjugate</i> | Enzyme Conjugate | Enzymkonjugat | Conjugué enzymatique | Conjugado enzimático | Tracciante enzimatico |
| <i>Enzyme Complex</i> | Enzyme Complex | Enzymkomplex | Complexe enzymatique | Complex enzimático | Complesso enzimatico |
| <i>Substrate Solution</i> | Substrate Solution | Substratlösung | Solution substrat | Solución de sustrato | Soluzione di substrato |
| <i>Zero Standard</i> | Zero Standard | Nullstandard | Standard 0 | Estándar 0 | Standard zero |
| <i>Standard</i> | Standard | Standard | Standard | Estándar | Standard |
| <i>Control</i> | Control | Kontrolle | Contrôle | Control | Controllo |
| <i>Substrat Solution Reagent A</i> | Substrat Solution Reagent A | Substratlösung Reagenz A | | | |
| <i>Substrat Solution Reagent B</i> | Substrat Solution Reagent B | Substratlösung Reagenz B | | | |
| <i>Substrat Solution Reagent C</i> | Substrat Solution Reagent C | Substratlösung Reagenz C | | | |
| <i>Assay Buffer</i> | Assay Buffer | Assaypuffer | Tampon d'essai | Tampón de ensayo | Tampone del test |
| <i>Wash Solution</i> | Wash Solution | Waschlösung | Solution de lavage | Solución de lavado | Soluzione di lavaggio |
| <i>1N NaOH</i> | 1N NaOH | 1N NaOH | 1N NaOH | 1N NaOH | 1N NaOH (idrossido di sodio 1N) |
| <i>1 N HCl</i> | 1 N HCl | 1 N HCl | 1N HCl | 1 N HCl | 1 N HCl (acido cloridrico 1N) |
| <i>Sample Diluent</i> | Sample Diluent | Probenverdünnungsmedium | Solution pour dilution de l'échantillon | Solución para dilución de la muestra | Diluyente dei campioni |
| <i>Conjugate Diluent</i> | Conjugate Diluent | Konjugatverdünnungsmedium | Solution pour dilution du conjugué | Solución para dilución del conjugado | Diluyente del tracciante |

| Symbol | Portugues | Dansk | Svenska | Ελληνικά |
|---|--------------------------------------|--|---|---------------------------------------|
|  | Conformidade com as normas europeias | Europæisk overensstemmelse | Europeisk överensstämmelse | Ευρωπαϊκή Συμμόρφωση |
|  | Instruções de uso | Brugermanual | Användar manual | Εγχειρίδιο χρήστη |
| IVD | Diagnóstico in vitro | In vitro diagnostik | Diagnostik in vitro | in vitro διαγνωστικό |
| RUO | | | | |
| REF | Catálogo n.º | Katalognummer | Katalog nummer | Αριθμός καταλόγου |
| LOT | No do lote | Lot nummer | Batch-nummer | Αριθμός Παρτίδος |
|  | | Indeholder tilstrækkeligt til "n" test | Innehåller tillräckligt till "n" tester | Περιεχόμενο επαρκές για «n» εξετάσεις |
|  | Temperatura de conservação | Opbevaringstemperatur | Förvaringstemperatur | Θερμοκρασία αποθήκευσης |
|  | Prazo de validade | Udløbsdato | Bäst före datum | Ημερομηνία λήξης |
|  | Fabricante | Producent | Tillverkare | Κατασκευαστής |
| <i>Distributed by</i> | | | | |
| <i>Content</i> | Conteúdo | Indhold | Innehåll | Περιεχόμενο |
| <i>Volume/No.</i> | Volume/Número | Volumen/antal | Volym/antal | Όγκος/αριθ.. |
| <i>Microtiterwells</i> | Alvéolos de microtitulação | Mikrotiterbrønde | Brunnar i Mikrotiterplatta | Πηγαδάκια Μικροπιλοδοτήσεως |
| <i>Antiserum</i> | Anti-soro | Antiserum | Antiserum | Αντιορός |
| <i>Enzyme Conjugate</i> | Conjugado enzimático | Enzymkonjugat | Enzymkonjugat | Συζευγμένο ενζυμο |
| <i>Enzyme Complex</i> | Complexo enzimático | Enzymkompleks | Enzymkomplex | Σύμπλοκο ενζύμου |
| <i>Substrate Solution</i> | Solução de substrato | Substratopløsning | Substratlösning | Διάλυμα υποστρώματος |
| <i>Zero Standard</i> | Padrão zero | Standard 0 | Standard 0 | Πρότυπο Μηδέν |
| <i>Standard</i> | Calibrador | Standard | Standard | Πρότυπα |
| <i>Control</i> | Controlo | Kontrol | Kontroll | Έλεγχος |
| <i>Substrat Solution Reagent A</i> | | | | |
| <i>Substrat Solution Reagent B</i> | | | | |
| <i>Substrat Solution Reagent C</i> | | | | |
| <i>Assay Buffer</i> | Tampão de teste | Assay buffer | Assay Buffer | Ρυθμιστικό Διάλυμα Εξέτασης |
| <i>Wash Solution</i> | Solução de lavagem | Vaskebuffer | Tvätt lösning | Διάλυμα πλύσεως |
| <i>1N NaOH</i> | 1N NaOH | 1N NaOH | 1N NaOH | 1N NaOH |
| <i>1 N HCl</i> | 1 N HCl | 1 N HCl | 1 N HCl | 1 N HCl |
| <i>Sample Diluent</i> | | | | |
| <i>Conjugate Diluent</i> | | | | |